

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------

TECHNOZYM® ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT71B0.01
 30-Apr-2019

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B1	30-Apr-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.42	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY51B1)	54.32 – 81.48 U/mL	67.13	OK
	Low Control (RPZ51B1)	5.27 – 12.29 U/mL	10.14	OK

Durchführung von
Prepared by

Michael KAFKA 

19. JUNI 2017
Datum / Date:

Freigabe von
Approved by

 Lucas ROSENMAYR

19. JUNI 2017
Datum / Date: