

**CERTIFICATE OF ANALYSIS**

# ***CHROMOGENIX***

PRODUCT: <u>FXa, 10 x 71 nkat</u>	PRODUCT # : <u>82098539</u>
EXP. DATE: <u>2019-01</u>	LOT # : <u>N0763936</u>
VENDOR: <u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM: <u>650515</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>FACTOR Xa, 71 nkat (650515)</b>					
ACTIVITY	61	81	nkat/vial	70	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	67.5	74.6	nkat/vial	68.6	PASS

**KIT COMPONENTS :**

NAME	LOT NUMBER	EXP.DATE	VOLUME
FXa, 71 nkat	N0763936	2019-01	2 ml

PREPARED BY: *N. Patel* DATE: Aug.3, 2016

APPROVAL SIGNATURE: *D. Tencza* DATE: Aug. 3, 2016