



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7C00AZ

TECHNOTHROMBIN[®] TGA RB 5x0.5mL

REF 5006209

LOT 7C71B00.01

30-Jun-2019

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RB	7C71B00	30-Jun-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	162 - 246	217.1	OK

Durchführung von
Prepared by

Fikret NASUFI

13. JULI 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA

13. JULI 2017

Datum / Date: