

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7F00AZ
---	---	----------

TECHNOTHROMBIN® TGA RC High 5x0.5mL

REF 5006214
LOT 7F61B00.03
 28-Feb-2018

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC High	7F61B00	28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	407 - 678	594.0	OK
	TGA Control Low (nM Peak Thrombin) LOT 2863B00	63 - 147	106.4	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

27. JUNI 2017
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

27. JUNI 2017
Datum / Date: