

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ8F00AZ
-----------------------------------------------------------------------------------	-------------------------------------------------------	----------

## TECHNOTHROMBIN<sup>®</sup> TGA SUB 5x1.5mL

**REF**      5006235  
**LOT**      8F63B00.02  
      31-Oct-2018

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
5x1.5mL	TGA Substrat	8F63B00	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <b>LOT</b> 1Z61B00	TGA RC High <b>LOT</b> 7F61B00	407 - 678	480.7    OK
		TGA RD <b>LOT</b> 7G61B00	402 - 670	481.9    OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

13. JULI 2017

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA *Michael Kafka*

13. JULI 2017

Datum / Date: