
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1P00.01
---	---	-----------

Coagulation Control N 50x1mL

REF 5020050
LOT 1P71B00.01
 31-May-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Siron LIS	26.0 – 32.0 sec	27.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Suzana GRESAK PODAKAR


20. JULI 2017

Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

20. JULI 2017

Datum / Date: