
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3C00.02
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------

TECHNOVIEW UFH Control L 5x1mL

REF 5090072
LOT 3C12B03.01
 31-Jan-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HbsAg HIV 1.2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® Alpha	Anti-Xa UFH	0.24 – 0.44 IU/mL	0.32	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Monika WEINBAUER

11. MRZ. 2016

Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

11. MRZ. 2016

Datum / Date: