



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3J00.02

TECHNOVIEW UFH CONT HIGH 5x1mL

REF 5090074
LOT 3J12B03.01
⌚ 30-Jun-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti-Xa (UFH)	0.66 – 1.10 U/mL	0.88	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa (UFH)	0.66 – 1.10 U/mL	0.86	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

30. AUG. 2016
Datum / Date:

Freigabe von
Approved by


Lavdrin NASUFI

31. AUG. 2016
Datum / Date: