
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3500AZ.01
---	---	-------------

TECHNOVIEW Rivaroxaban CONT M 5x1mL

REF 5090173
LOT 3551B01.01
 30-Apr-2019

Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron Alpha®	Anti Xa Rivaroxaban	114.9 – 191.5 ng/mL	144.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas ROSENMAYR

23. JUNI 2017

Datum / Date:

**Freigabe von
Approved by**

Stefan STEINER

23. JUNI 2017

Datum / Date: