



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1R00.01

Coagulation Reference, 5x1mL

REF 5220110  
LOT 1R71B00.01  
⌚ 28-Feb-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen Clauss method, diluted plasma	2.70 – 4.06 g/L	3.39	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Lucas ROSENMAYR

12. JULI 2017

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA

13. JULI 2017

Datum / Date: