





ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0J00AZ.01


TECHNOCHROM® C1-Inhibitor 30T.

REF 5345003
LOT 0J72B00.01
 30-Sep-2018

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|----------------------------------|------------|---|
| 1x1mL | Coagulation Control N for C1-INH | 1P72BC1H | 28-Feb-2019 |
| 1x1mL | Coagulation Control A for C1-INH | 3P73BC1H | 31-Mar-2019 |
| 1x0.982mL | Coagulation Reference for C1-INH | 1R72BC1H | 31-Mar-2019 |
| 1x3mL | C1 Esterase | 7B64B00 | 30-Sep-2018 |
| 1x3mL | Substrate f. C1-Esterase | 8C71B00 | 31-Jan-2022 |
| 1x25mL | C1-INH Probenpuffer A | 9M62B00 | 31-Dec-2021 |
| 1x20mL | C1-INH Reaktionspuffer B | 9N62B00 | 31-Dec-2021 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|----------------------------|----------------------------|--------------------|----|
| Verwendbarkeit Usability Ceveron® alpha | CCN LOT 1P72BC1H | 88.1 – 146.9% | 117.0 | OK |
| | CCA LOT 3P73BC1H | 41.7 – 69.5% | 54.7 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

Durchführung von
Prepared by

Suzana GRESAK PODAKAR


23. MAI 2017

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

23. MAI 2017

Datum / Date: