




ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ4C00AZ

Protein C Mangelplasma 2x1 mL

REF 5346806

LOT 4C42B01.04

 31-Dec-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.7	OK
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	CCN LOT 1P37B01	69.8 – 116.3%	105.8 OK
	CCA LOT 3P38B01	13.5 – 22.5%	20.4 OK
Etiketten Überprüfung Label Verification		OK	

Durchführung von  
Prepared by

  
Christoph ENNE

07. DEZ. 2016

Datum / Date:

Freigabe von  
Approved by

  
Lucas ROSENMAYR

07. DEZ. 2016

Datum / Date: