



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6NG00AZ

TECHNOZYM® vWF:Ag Calibrator Set 5x0.5mL

REF 5450210

LOT 6NG71B0.01

⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
OD	Calibrator 1	> 1.0	1.2	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.44 – 0.74 IU/mL	0.66	OK
	Low Control (BND61B0)	0.10 – 0.22 IU/mL	0.19	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

06. JULI 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

06. JULI 2017

Datum / Date: