

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7NG00AZ
---	---	-----------

TECHNOZYM® vWF:Ag Control Set 2x0.5mL

REF 5450212
LOT 7NG71B0.01
 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High Control (AND61B0)	0.44 – 0.74 IU/mL	0.56	OK
	Low Control (BND61B0)	0.10 – 0.22 IU/mL	0.18	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

06. JULI 2017
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

06. JULI 2017
Datum / Date: