



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM® vWF:CBA ELISA Kit 96T

REF 5450301

LOT RB75B00.01

⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	vWF:CBA Plate Collagen Typ III	RNC72B0	30-Apr-2019
1 x 0.3mL	CBA Conjugate	RNE71B0	30-Apr-2019
1x 80mL	Wash buffer Concentrate	RNA64B0	31-Dec-2018
1 x 90mL	Incubation buffer	RNB71B0	31-Jan-2019
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK71B0	31-May-2019
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.39	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.63 – 1.05 IU/mL	0.89	OK
	Low Control (BND61B0)	0.13 – 0.29 IU/mL	0.19	OK

Durchführung von  
Prepared by

Michael KAFKA 

13. JUNI 2017  
Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

13. JUNI 2017  
Datum / Date: