

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401
LOT RP74B00.01
 30-Apr-2019

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 INH Plate	RPS75B0	31-May-2019
1 x 0.3mL	Adamts13 INH Conjugate	RPV71B0	30-Apr-2019
1x 80mL	Wash buffer Concentrate	RNA71B0	30-Apr-2019
1 x 90mL	Incubation buffer	RNB72B0	30-Apr-2019
1 x 12mL	TMB Substrate	RKN71B0	31-May-2019
1 x 12mL	Stop solution	RKK71B0	31-May-2019
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH High Control	RPY51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B1	30-Apr-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.32	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RPY51B1)	54.32 – 81.48 U/mL	70.81	OK
	Low Control (RPZ51B1)	5.27 – 12.29 U/mL	10.90	OK

Durchführung von
Prepared by

Michael KAFKA 

10. JULI 2017

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

11. JULI 2017

Datum / Date: