



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6T0A.02

Ceveron Technoplastin HIS 12x4mL

REF 5003010

LOT 6T5AB.01

⌚ 30-Nov-2017

Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Heparinempfindlichkeit Heparin Sensitivity	Abweichung zur 0 E/mL UFH Deviation to 0 E/mL UFH	1 E/mL UFH	± 10%	- 1.32	OK
Verwendbarkeit Technoplastin HIS Usability Technoplastin HIS Ceveron® alpha	Technoplastin HIS %	CCN LOT 1P37B00	109.7 – 138.3% .	128.2	OK
	Technoplastin HIS INR	CCN LOT 1P37B00	0.77 – 0.97	0.83	OK
	Technoplastin HIS %	CCA LOT 3P38B00	25.7 – 32.3%	28.0	OK
	Technoplastin HIS INR	CCA LOT 3P38B00	2.54 – 3.20	2.97	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Christoph Enne
Christoph ENNE

02. FEB. 2016

Datum / Date:

Freigabe von
Approved by

Lucas Rosenmayr
Lucas ROSENMAYR

02-Feb-2016

Datum / Date: