



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6T0AAZ.02

Ceveron® Technoplastin HIS 12x4mL

REF 5003010

LOT 6T69B.01

⌚ 31-Dec-2018

Analytische Verfahren Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Heparinempfindlichkeit Heparin Sensitivity	Abweichung zur 0 E/mL UFH Deviation to 0 E/mL UFH	1 E/mL UFH	± 10%	-4.28	OK
Verwendbarkeit FII Ceveron® alpha Usability FII Ceveron® alpha	CCN [LOT] 1P37B01		73.5 – 122.5 %	107.9	OK
	CCA [LOT] 3P38B01		12.8 – 21.3 %	18.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

14. APR. 2017

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

18. APR. 2017

Datum / Date: