



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ9P00AZ

Ceveron[®] TGA BUF 5x3 mL

REF 5006362

LOT 9P31B02.01

⌚ 31-Jul-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
5 x 3 mL	TGA buffer	9P31B02	31-Jul-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / usability am Ceveron [®] alpha	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	1.0 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

27. JUNI 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

27. JUNI 2017

Datum / Date: