

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0U00AZ.02
---	---	-------------

AK-Calibrant 4x1mL

REF 5010004
LOT 0U71B00.03
 31-Oct-2019

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	AK Calibrant A	1N64B01	31-Oct-2019
1x1mL	AK Calibrant B	3L64B01	31-Oct-2019
1x1mL	AK Calibrant C	3M6DB01	31-Oct-2019
1x1mL	AK Calibrant D	3N68B01	31-Oct-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TCN LOT 1A71B00 Technoclot PT Plus	68.9 – 86.7 %	75.2	OK
	CCA LOT 3P38B01 Technoclot PT Plus	12.4 – 15.6 %	14.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

17. JULI 2017
Datum / Date:

**Freigabe von
Approved by**


Suzana GRESAK PODAKAR

17. JULI 2017
Datum / Date: