




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6S0AAZ.01

Siron LS (aPTT liquid) 2x4 mL

REF 5035105

LOT 6S68BA0.01

 31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.6	7.0	OK
Sterilität sterility		0 CFU/mL	0	OK
Sensitivität Lupus sensitivity lupus	Lupus Inhibitor Plasma	≥ 1.5	3.12	OK
aPTT Verwendbarkeit aPTT Usability Ceveron® alpha	CCN LOT 1P37B01	23.6 – 32.0 sec	28.0	OK
	CCA LOT 3P38B01	59.2 – 80.0 sec	69.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Christoph ENNE

19. DEZ. 2016

Datum / Date:

Freigabe von
Approved by


Stefan STEINER

19. DEZ. 2016

Datum / Date: