

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6E0A.01
---	---	-----------

Fibrinogen Reagenz 5x2mL

REF 5138085
LOT 6E68BA0.02
 31-Okt-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.0 – 8.0	7.8	OK
Restfeuchte Residual moisture	≤3%	1.61	OK

Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B01	2.52 – 3.78 g/L	3.41	OK
	CCA LOT 3P38B01	0.93 – 1.39 g/L	1.07	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas
Lucas ROSENMAYR

07. JUNI 2017

Datum / Date:

Freigabe von
Approved by

Stefan
Stefan STEINER

07. JUNI 2017

Datum / Date: