

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0N00.02
---	---	-----------


TECHNOCHROM® F VIII:C 40T.

REF 5344101
LOT 0N71B00.01
 30-Apr-2019

Inhalt Content	Reagenz Reagent	LOT	
2x2mL	Substrate for factor Xa 10µmol	8M62B00	30-Sep-2021
2x2mL	F VIII:C Reagenz A	7M61B01	30-Apr-2019
2x2mL	F VIII:C Reagenz B	7L41B02	30-Apr-2019
1x1mL	Reference Standard 1, F VIII:C	2W61B01	30-Apr-2019
1x1mL	Reference Standard 2, F VIII:C	2X61B01	30-Apr-2019
1x1mL	Reference Standard 3, F VIII:C	2Y41B03	30-Apr-2019
1x1mL	Reference Standard 4, F VIII:C	2Z51B0E	30-Apr-2019
2x8mL	FVIII:C Reaction buffer	9T63B00	31-Dez-2021
2x30mL	FVIII:C Dilution buffer	9Y61B00	31-Dez-2021

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B01	69.0 – 115.0%	91.5	OK
	CCA LOT 3P38B01	19.5 – 32.5%	28.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lavdrim NASUFI

16. JUNI 2017

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

16. JUNI 2017

Datum / Date: