



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3V00AZ.01

TECHNOVIEW Arixtra CONT H 6x1mL

REF 5090014

LOT 3V51B01.01

🕒 31-Mai-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.62 – 1.04 µg/mL	0.83	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.62 – 1.04 µg/mL	0.85	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

19. JUNI 2017
Datum / Date:

Freigabe von
Approved by


Christoph ENNE

19. JUNI 2017
Datum / Date: