



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0900AZ.01

TECHNOVIEW® Rivaroxaban Cal Set 5x1mL

REF 5090170
LOT 0961B00.03
⌚ 30-Jun-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 1	5103B02	30-Jun-2018
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 2	5601B02	30-Jun-2018
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 3	5711B02	30-Jun-2018
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 4	5811B02	30-Jun-2018
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 5	5952B01	30-Jun-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Anti-Xa Rivaroxaban Ceveron® alpha	Rivaroxaban Cont M LOT 3551B00 114.9 – 191.5 ng/mL	150.1 OK
	Rivaroxaban Cont L LOT 3453B00 41.6 – 77.2 ng/mL	56.0 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by


Lucas ROSENMAYR

05. MAI 2017

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

05. MAI 2017

Datum / Date: