

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0300.01
---	---	-----------

Technoview® Rivaroxaban CAL Set High 5x1mL

REF 5090171
LOT 0371B00.01
 31-May-2019

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Technoview Rivaroxaban CAL1 High	5161B01	31-May-2019
1x1mL	Technoview Rivaroxaban CAL2 High	5261B01	31-May-2019
1x1mL	Technoview Rivaroxaban CAL3 High	5361B01	31-May-2019
1x1mL	Technoview Rivaroxaban CAL4 High	5411B04	31-May-2019
1x1mL	Technoview Rivaroxaban CAL5 High	5561B01	31-May-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control High LOT 3351B01	242.6 – 404.3 ng/mL	356.2	OK
	Rivaroxaban Control Medium LOT 3551B01	114.9 – 191.5 ng/mL	168.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Christoph ENNE

30. AUG. 2017
Datum / Date:

Freigabe von
Approved by


Suzana GRESAK PODAKAR

31. AUG. 2017
Datum / Date: