
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2K00.02
---	---	-----------

FVIII Inhibitorplasma 5x1mL

REF 5159008
LOT 2K52B01.02
 31-Jan-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	FVIII Inhibitor clotting, Daptin TC	13.5 – 22.5 BE/mL	22.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas Rosenmayr
Lucas ROSENMAYR

13. JULI 2017

Datum / Date:

**Freigabe von
Approved by**

Christoph Enne
Christoph ENNE

13. JULI 2017

Datum / Date: