



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2V00.02

Lupus Inhibitor Plasma 2x1mL

REF 5343010

LOT 2V62B00.01

⌚ 31-Jul-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.7	OK
Restfeuchte Residual moisture	≤1%	0.31	OK
OD _{320nm}	≤10	5.5	OK
Verwendbarkeit Usability Ceveron® alpha	LA Ratio LA Screen/LA Confirm	1.45 – 2.41	2.20 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

28. FEB. 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

28. FEB. 2017

Datum / Date: