

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00AZ.01
---	---	-------------

TECHNOCHROM® C1-Inhibitor 30T.

REF 5345003
LOT 0J73B00.02
 28-Feb-2019

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N for C1-INH	1P72BC1H	28-Feb-2019
1x1mL	Coagulation Control A for C1-INH	3P73BC1H	31-Mar-2019
1x0.982mL	Coagulation Reference for C1-INH	1R72BC1H	31-Mar-2019
1x3mL	C1 Esterase	7B71B00	30-Apr-2019
1x3mL	Substrate f. C1-Esterase	8C71B00	31-Jan-2022
1x25mL	C1-INH Probenpuffer A	9M62B00	31-Dec-2021
1x20mL	C1-INH Reaktionspuffer B	9N62B00	31-Dec-2021

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P72BC1H	88.1 – 146.9%	118.3	OK
	CCA LOT 3P73BC1H	41.7 – 69.5%	51.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

25. AUG. 2017

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

25. AUG. 2017

Datum / Date: