

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6ND00AZ
---	---	-----------

TECHNOZYM® vWF:CBA Calibrator Set 5x0.5mL

REF 5450310
LOT 6ND61B0.01
 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.40	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.63 – 1.05 IU/mL	0.84	OK
	Low Control (BND61B0)	0.13 – 0.29 IU/mL	0.24	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

08-Jun-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

08-Jun-2016
Datum / Date: