



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRQ00AZ

TECHNOZYM® ADAMTS-13 ELISA Kit 2x48T.

REF 5450501

LOT RQ71B00.01

30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	
2x48T	Adamts13 Plate	RQS61B0	30-Apr-2018
1x0.3mL	Adamts13 Conjugate	RQV32B2	30-Apr-2018
2x3mL	AdamTS13 Activity Substrate	RQW61B0	30-Apr-2018
1x6mL	AdamTS13 Antigen Substrate	RQX61B0	30-Jun-2018
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL61B0	30-Jun-2018
1x6mL	AdamTS13 Antigen Stop Solution	RQK61B0	30-Jun-2018
1x80mL	Wash buffer Concentrate	RNA63B0	31-Aug-2018
1x90mL	Incubations buffer	RNB63B0	31-Aug-2018
1x0.5mL	Adamts 13 Calibrator 1	1RT61B0	30-Nov-2018
1x0.5mL	Adamts 13 Calibrator 2	2RT61B0	30-Nov-2018
1x0.5mL	Adamts 13 Calibrator 3	3RT61B0	30-Nov-2018
1x0.5mL	Adamts 13 Calibrator 4	4RT61B0	30-Nov-2018
1x0.5mL	Adamts 13 Calibrator 5	5RT61B0	31-Dec-2018
1x0.5mL	Adamts 13 High Control	RRY61B0	31-Dec-2018
1x0.5mL	Adamts 13 Low Control	RRZ61B0	31-Dec-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RRY61B0 activity	0.46 – 0.76 IU/mL	0.58	OK
	High control RRY61B0 antigen	0.54 – 0.90 IU/mL	0.74	OK
	Low control RRZ61B0 activity	0.11 – 0.25 IU/mL	0.16	OK
	Low control RRZ61B0 antigen	0.14 – 0.32 IU/mL	0.22	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	≥200	411.4	OK
	Gain: 59 Antigen ΔRFU Cal1-Cal5	≥200	363	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQRQ00AZ</p>
---	---	---

Durchführung von
Prepared by

Fikret NASUF *Fikret Nasuf*

08. FEB. 2017

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR *Lucas*

08. FEB. 2017

Datum / Date: