

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ8UT0AZ.02
---	---	-------------

**TECHNOZYM® ADAMTS-13 Activity Calibrator Set 6x0.5mL**

**REF**            5450761  
**LOT**            8UT71B0.02  
        31-Oct-2018

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x0.5mL	Adamts-13 Activity Calibrator 1	1UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 2	2UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 3	3UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 4	4UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 5	5UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 6	6UT63B0	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.41	OK
	Calibrator 6	< 0.2	0.02	OK
Verwendbarkeit / Usability	High Control (RUY63B0)	0.59 – 0.98 IU/mL	0.83	OK
	Low Control (RUZ63B0)	0.16 – 0.38 IU/mL	0.26	OK


**Durchführung von  
Prepared by**

Michael KAFKA 

05. JULI 2017

Datum / Date:

**Freigabe von  
Approved by**

Fikret NASUFI 

05. JULI 2017

Datum / Date: