

CERTIFICATE OF ANALYSIS

# CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 1	PRODUCT #:	82265063
EXP. DATE:	2019-06	LOT #:	N0570180
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	C00423

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION	
	MIN	MAX				
<b>Analysis with COATEST APC R:</b>						
APT Time	CaCl <sub>2</sub>	28	34	seconds	30	PASS
	APC/CaCl <sub>2</sub>	90	--	seconds	116	PASS
APC Ratio		3.5	4.4	--	3.8	PASS

<b>Analysis with COATEST APC R V:</b>						
APT Time	CaCl <sub>2</sub>	--	38	seconds	34	PASS
	APC/CaCl <sub>2</sub>	75	--	seconds	105	PASS
APC Ratio		2.7	3.2	--	3.1	PASS

**Kit Components:**

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 1	N0570180	2019-06	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel N.M. Patel

DATE: 6/15/2017 6.15.17

APPROVAL SIGNATURE: D. Tencza P. Tencza

DATE: 6/15/2017 6/15/17