

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2302, 25 mg Substrate	PRODUCT #: 00082034039
EXP. DATE: 2020-02	LOT #: N1065926
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N1065926	2020-02	3 mL

PREPARED BY:

Maia Dettinger

DATE:

Nov 7, 2016

APPROVAL SIGNATURE:

J. Penza

DATE:

11/14/16

Form QCM00820340-2 Rev. 03/09