

# TECO medical

## TECO® Perch Vitellogenin ELISA KIT

Certificate of Analysis

Analysezertifikat

Certificat d'Analyse

**REF**  
**LOT**

TE1035  
050717

**EXP**

2018-05-06

Refer to Product Insert for further Information  
Weitere Informationen hierzu finden Sie in der Packungsbeilage des Produkts  
Se référer à la Notice produit pour plus d'informations

Standard Curve

Eichkurve

Courbe Standard

Standards	Target Values / Zielwerte / objectif		Extinction at 450 nm
A	80	ng/ml	2,528
B	27	ng/ml	1,090
C	9	ng/ml	0,468
D	3	ng/ml	0,229
E	1	ng/ml	0,147
F	0	ng/ml	0,107
L - Control 1	5,2 (3,4 - 7,0)		0,321
H - Control 2	24,4 (15,9 - 32,9)		1,014

**ITEM**

**LOT**

**EXP**

① pVTG specific antibody coated wells	PS070617	2018-09-30
② Wash Buffer 50x	WA840717	2019-06-30
③ Dilution Buffer	DO140717	2018-05-06
④ Matrix Solution	AR050717	2018-12-31
⑤ Biotinylated Antibody	BN070717	2018-12-31
⑥ SA-HRP Conj.	EZ010717	2018-06-30
⑦ TMB Substrate	TA430417	2019-03-21
⑧ Stop Solution	SA640717	2019-06-30
S Standard Stock 10x	ZG060717	2018-06-30
L Control 1 10x	ZG060717	2018-06-30
H Control 2 10x	ZG060717	2018-06-30

TECOmedical

Qc release

Date

2017-07-12