



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0T00AZ

TECHNOTHROMBIN[®] TGA Kit 3x16 T.


REF 5006010

LOT 0T71B00.01

⌚ 31-Mar-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x3mL	TGA buffer	9P31B02	31-Jul-2018
3x1.5mL	TGA substrat	8F63B00	31-Oct-2018
1x1.5mL	TGA reagent RD	7G61B00	30-Jun-2018
1x0.5mL	TGA reagent RC Low	7D62B00	31-May-2018
1x0.5mL	TGA reagent RC High	7F72B00	31-Aug-2019
1x0.5mL	TGA Thrombin Calibrator	5T61B	31-Mar-2018
1x1mL	TGA Control High	1Z71B00	30-Jun-2018
1x1mL	TGA Control Low	2863B01	30-Apr-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1Z71B00	336 - 560	474.0	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2863B01	63 - 147	127.8	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1Z71B00	207 - 345	292.4	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2863B01	57 - 143	125.5	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1Z71B00	369 - 615	523.7	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2863B01	365 - 608	515.7	OK

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ0T00AZ</p>
---	---	---

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

20. SEP. 2017

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR
S. Gresak Podakar

20. SEP. 2017

Datum / Date: