

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7F00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA RC High 5x0.5mL

REF      5006214  
LOT      7F72B00.01  
      31-Aug-2019

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x0.5 mL	TGA RC High	7F72B00	31-Aug-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z71B00	336 - 560	484.2	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2863B01	63 - 147	127.8	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

20. SEP. 2017  
Datum / Date:

**Freigabe von**  
**Approved by**

*Christoph Enne*  
Christoph ENNE

20. SEP. 2017  
Datum / Date: