
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7Q00AZ
---	---	----------

Ceveron[®] TGA RC High 5x1 mL

REF 5006227
LOT 7Q71B.02
 31-Jul-2019

Inhalt Content	Reagenz Reagent	LOT	
5x1 mL	Ceveron [®] TGA RC High	7Q71B	31-Jul-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Y71B00	267 - 445	330.4	OK
	TGA Control Low (nM Peak Thrombin) LOT 2772B01	83 - 139	121.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19. SEP. 2017
Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR
S. Gresak Podakar

20. SEP. 2017
Datum / Date: