

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0D00AZ
---	---	----------


## TECHNOTHROMBIN® TGA Cal Set

REF      5006345  
LOT      0D63B00.02  
      31-Mar-2018

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 3 mL	TGA buffer	9P31B02	31-Jul-2018
1 x 0.5mL	TGA Thrombin Calibrator	5T61B	31-Mar-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580		800-900 nM	841	OK
Verwendbarkeit / usability  Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	0.99	OK

**Durchführung von  
Prepared by**

Fikret NASUFI 

20. SEP. 2017

Datum / Date:

**Freigabe von  
Approved by**

  
Lucas ROSENMAYR

20. SEP. 2017

Datum / Date: