




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQ1P00AZ.01

Coagulation Control N 50x1mL

REF 5020050


LOT 1P71B00.02

 31-May-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Technoplastin® HIS – Cev (%)	96.7 – 121.9	111.2	OK
	Technoplastin® HIS – Cev (INR)	0.80 – 1.00	0.91	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von
Prepared by**


Laydjin NASUEI

03. OKT. 2017

Datum / Date:

**Freigabe von
Approved by**


Stefan STEINER

03. OKT. 2017

Datum / Date: