



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3600.01

TECHNOVIEW® Dabigatran CONT High 5x1mL

REF 5090212
LOT 3642B02.01
⌚ 31-Mai-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		236.7 – 394.5 ng/mL	315.6	OK
Verwendbarkeit Usability Ceveron® alpha	Technoclot DTI	236.7 – 394.5 ng/mL	292.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

29. SEP. 2017

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

29. SEP. 2017

Datum / Date: