



| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ3700.01 |
|---|---|-----------|

TECHNOVIEW Dabigatran CONT Low 5x1mL

REF 5090214
LOT 3742B02.01
 31-Mai-2019

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|---|---|----------------------------------|----------------------------------|----|
| pH | | 7.4 – 7.8 | 7.6 | OK |
| Virologie Virology | HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab | Negative Negative Negative | Negative Negative Negative | OK |
| Chargenspezifische Kalibrationswerte Lot specific calibration values | | 98.9 – 183.7 ng/mL | 141.3 | OK |
| Verwendbarkeit Usability Ceveron® alpha | Technoclot DTI | 98.9 – 183.7 ng/mL | 114.0 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

Durchführung von
Prepared by

Lucas
Lucas ROSENMAYR

29. SEP. 2017

Datum / Date:

Freigabe von
Approved by

Suzana
Suzana GRESAK PODAKAR

29. SEP. 2017

Datum / Date: