



	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6ND00AZ
---	---	-----------

TECHNOZYM® vWF:CBA Control Set 2x0.5mL

REF 5450312
LOT 7ND61B0.01
 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High Control (AND61B0)	0.63 – 1.05 IU/mL	0.84 OK
	Low Control (BND61B0)	0.13 – 0.29 IU/mL	0.24 OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

08-Jun-2016
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *Michael Kafka*

08-Jun-2016
Datum / Date: