



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401

LOT RP76B00.01

🕒 30-Apr-2019

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 12	Adamts13 INH Plate	RPS76B0	31-Jul-2019
1 x 0.3mL	Adamts13 INH Conjugate	RPV71B0	30-Apr-2019
1x 80mL	Wash buffer Concentrate	RNA71B0	30-Apr-2019
1 x 90mL	Incubation buffer	RNB72B0	30-Apr-2019
1 x 12mL	TMB Substrate	RKN71B0	31-May-2019
1 x 12mL	Stop solution	RKK71B0	31-May-2019
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	31-Aug-2019
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	31-Aug-2019
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	31-Aug-2019
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	31-Aug-2019
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	31-Aug-2019
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	31-Aug-2019
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	31-Aug-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.31	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	77.35	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	28.71	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

18. SEP. 2017

Datum / Date:

Freigabe von  
Approved by

Lucas ROSENMAYR *Lucas Rosenmayr*

18. SEP. 2017

Datum / Date: