



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451
LOT RP75B0S.01
⌚ 30-Nov-2018

| Inhalt Content | Reagenz Reagent | LOT | ⌚ |
|-------------------|---------------------------|---------|-------------|
| 1 x 6 | Adamts13 INH Plate | RPS6BBS | 30-Nov-2018 |
| 1 x 0.3mL | Adamts13 INH Conjugate | RPV71B0 | 30-Apr-2019 |
| 1x 80mL | Wash buffer Concentrate | RNA64B0 | 31-Dec-2018 |
| 1 x 90mL | Incubation buffer | RNB71B0 | 31-Jan-2019 |
| 1 x 12mL | TMB Substrate | RKN71B0 | 31-Mai-2019 |
| 1 x 12mL | Stop solution | RKK71B0 | 31-Mai-2019 |
| 1 x 0.5mL | Adamts13 INH Calibrator 1 | 1PT71B0 | 31-Aug-2019 |
| 1 x 0.5mL | Adamts13 INH Calibrator 2 | 2PT71B0 | 31-Aug-2019 |
| 1 x 0.5mL | Adamts13 INH Calibrator 3 | 3PT71B0 | 31-Aug-2019 |
| 1 x 0.5mL | Adamts13 INH Calibrator 4 | 4PT71B0 | 31-Aug-2019 |
| 1 x 0.5mL | Adamts13 INH Calibrator 5 | 5PT71B0 | 31-Aug-2019 |
| 1 x 0.5mL | Adamts13 INH High Control | RPY71B0 | 31-Aug-2019 |
| 1 x 0.5mL | Adamts13 INH Low Control | RPZ71B0 | 31-Aug-2019 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | > 1.0 | 1.44 | OK |
| | Calibrator 5 | < 0.1 | 0.00 | OK |
| Verwendbarkeit Usability | High Control (RPY71B0) | 60.24 – 100.40 U/mL | 81.36 | OK |
| | Low Control (RPZ71B0) | 16.64 – 38.84 U/mL | 25.23 | OK |

Durchführung von
Prepared by

Fikret NASUFI

10. SEP. 2017

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR

10. SEP. 2017

Datum / Date: