



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM<sup>®</sup> ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601

LOT RR73B00.01

⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 Ag Plate	RRS72B0	30-Jun-2019
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B2	30-Apr-2018
1x 80mL	Wash buffer Concentrate	RNA71B0	30-Apr-2019
1 x 90mL	Incubation buffer	RNB72B0	30-Apr-2019
1 x 12mL	TMB Substrate	RKN71B0	31-May-2019
1 x 12mL	Stop solution	RKK71B0	31-May-2019
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT61B0	31-Dec-2018
1 x 0.5mL	Adamts13 Ag High Control	RRY61B0	31-Dec-2018
1 x 0.5mL	Adamts13 Ag Low Control	RRZ61B0	31-Dec-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.81	OK
Verwendbarkeit Usability	High Control (RRY61B0)	0.47 – 0.79 IU/mL	0.63	OK
	Low Control (RRZ61B0)	0.13 – 0.29 IU/mL	0.19	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

01. AUG. 2017  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *[Signature]*

01. AUG. 2017  
Datum / Date: