

CHROMOGENIX

Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2019-06	LOT #: N0771002
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0090	PASS
Intercept	--	0.250	--	0.129	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.046	PASS
Intercept	--	0.406	--	0.174	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0670770	2020-06	(1x) 6 ml
CT SP Substrate	N0670772	2019-06	(1x) 12 ml
Factor Reagent	N0972284	2020-03	(4x) 3 ml
CT SP Buffer, pH 7.3	N0670624	2019-06	(1x) 20 ml
CT SP Phospholipid	N0670769	2019-06	(1x) 2 ml

PREPARED BY: MOETTMGER *M. Moettinger* DATE: 09/28/17 *9-28-17*

APPROVAL SIGNATURE: PATEL *P. Patel* DATE: 09/28/17 *9-28-17*