



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ...00AZ

TECHNOZYM® PAP Complex ELISA Kit 96T.

REF TC12060

LOT 120602B7.01

⌚ 30-Jun-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	PAP Plate	RGY71B0	30-Apr-2019
1 x 0.3mL	anti PAP POX	RBK61B0	30-Jun-2018
2 x 20mL	Wash buffer Concentrate	RBE71B0	28-Feb-2019
1 x 20mL	Dilution Buffer	RAF61B0	30-Jun-2018
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK71B0	31-May-2019
1 x 0.5mL	PAP-Calibrator 1	1BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 2	2BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 3	3BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 4	4BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 5	5BM61B0	30-Jun-2018
1 x 0.5mL	PAP-high Control	ABM61B0	31-Jul-2018
1 x 0.5mL	PAP-low Control	BBM61B0	31-Jul-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.24	OK
	Calibrator 5	OD: < 0.2	0.02	OK
Verwendbarkeit Usability	High Control (ABM61B0)	98.9 – 164.9 ng/mL	130.82	OK
	Low Control (BBM61B0)	13.0 – 30.2 ng/mL	21.83	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

27. SEP. 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

27. SEP. 2017

Datum / Date: