

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
-----------------------------------------------------------------------------------	-------------------------------------------------------------	----------

TECHNOTHROMBIN® TGA RC Low 5x0.5mL

REF 5006212
LOT 7D62B00.03
 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC Low	7D62B00	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z71B00	207 - 345	284.8	OK
	TGA Control Low (nM Peak Thrombin) LOT 2863B01	57 - 143	118.2	OK

Durchführung von
Prepared by

Ehret Nasabi

20. SEP. 2017
Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

20. SEP. 2017
Datum / Date: