

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230
LOT 8F63B00.04
 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Substrat	8F63B00	31-Oct-2018

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z71B00	TGA RC High LOT 7F72B00	336 - 560	537.2	OK
		TGA RC Low LOT 7D62B00	207 - 345	300.1	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

20. SEP. 2017

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR *Lucas Rosenmayr*

20. SEP. 2017

Datum / Date: